

CRCST Practice Test Questions and Answers

1. What is the minimum temperature required for steam sterilization in an autoclave?

- A) 121°C (250°F)
- B) 100°C (212°F)
- C) 135°C (275°F)
- D) 160°C (320°F)

2. Which method is most effective for removing biofilm from surgical instruments?

- A) Manual cleaning with enzymatic detergent
- B) High-temperature washing only
- C) Ultrasonic cleaning with appropriate detergent
- D) Air drying at room temperature

3. What is the recommended relative humidity level in sterile storage areas?

- A) 30-40%
- B) 30-60%
- C) 60-80%
- D) 80-90%

4. Which personal protective equipment (PPE) is required when handling contaminated instruments?

- A) Gloves only
- B) Gown and gloves
- C) Full PPE including mask, gown, gloves, and eye protection
- D) Hair covering only

Answers: 1-A 2-C 3-B 4-C

For More CRCST Questions and Answers FREE, CRCST Online Prep Training, CRCST Exam, CRCST Study Guide, CRCST Flashcards, CRCST Quizzes visit:

CRCST Practice Test